

M74E

Rel. 1.03 of 19/09/14

Advanced DMM with safety and safety test verifies

Pag 1 of 3

1. TECHNICAL SPECIFICATIONS - DMM FUNCTIONS

Accuracy is indicated as \pm [% readings + (no. of digits*resolution)] at 23 °C \pm 5 °C, relative humidity <70%HR

DC/AC TRMS VOLTAGE (Autorange)							
Range	Resolution	DC Accuracy	Accuracy (30 ÷ 70Hz)	Accuracy (70 ÷ 400Hz)	Input impedance		
1.0 ÷ 999.9mV	0.1mV						
1.000 ÷ 9.999V	0.001V	1/0 E0/rda 10dat)	1/4 00/mdm 1 0dmt\	1/2 00/mdm 1 2 dmt)	1140		
10.00 ÷ 99.99V	0.01V	\pm (0.5%rdg+2dgt)	\pm (1.0%rdg+2dgt)	±(2.0%rdg+2dgt)	1ΜΩ		
100.0 ÷ 605.0V	0.1V						

AC/DC VOLTAGE: MAX / MIN / AVG / PEAK						
Function	Range	Resolution	Accuracy	Response time		
	1.0mV ÷ 999.9mV	0.1mV				
MAX, MIN,	1.000V ÷ 9.999V	1mV		500ms		
AVG	10.00V ÷ 99.99V	10mV	±(5.0%rdg + 10dgt)			
	100.0V ÷ 605.0V	100mV				
	10.0mV ÷ 999.9mV	0.1mV		1ms		
PEAK	1.000V ÷ 9.999V	1mV				
	10.00V ÷ 99.99V	10mV				
	100.0V ÷ 605.0V	100mV				

DC/AC CURRENT TRMS (with external clamp)						
Range	Resolution	DC Accuracy	Accuracy (30 ÷ 70Hz)	Accuracy (70 ÷ 400Hz)	Crest factor	Overload protection
1.0 ÷ 999.9mV	0.1mV	±(0.5%rdg+2	±(1.0%rdg+2	±(2.0%rdg+2	3	605\/rmc_may
1.000 ÷ 1.200V	1mV	dgt)	dgt)	dgt)	1.5	605Vrms max

Note: accuracy indicated don't consider clamp accuracy. Please refer also to transducers clamp user's manual.

AC/DC CURRENT: MAX / MIN / AVG / PEAK (with external clamp)						
Function	Range	Resolution	Accuracy	Respon se time	Overload protection	
MAX, MIN,	1.0mV ÷ 999.9mV	0.1mV		500 ms	605Vrms max	
AVG	1.000V ÷ 1.200V	1mV	1/F 00/rda 10 dat)	300 1118		
PEAK	10.0mV ÷ 999.9mV	0.1mV	±(5.0%rdg+10 dgt)		005VIIIS IIIAX	
FEAR	1.000V ÷ 3.000V	1mV		1ms		

RESISTANCE AND CONTINUITY TEST						
Range	Resolution	Accuracy	Continuity test	Overload protection		
$0.00\Omega \div 39.99\Omega$	0.01Ω					
$40.0\Omega \div 399.9\Omega$	0.1Ω	1/4 00/ md m 1 E d mt)	R ≤40Ω	605Vrms max for 1		
$400\Omega \div 3999\Omega$	1Ω	\pm (1.0%rdg+5 dgt)	R ≥40Ω	minute		
4.00 k $\Omega \div 39.99$ k Ω	10Ω					

FREQUENCY (with test leads)					
Range	Resolution	Accuracy	Input voltage	Overload protection	
30.0 ÷ 199.9Hz	0.1Hz	±(0 50/rda+2 dat)	1.0mV ÷ 605V	605Vrms max	
200 ÷ 400Hz	1Hz	\pm (0.5%rdg+2 dgt)	VC00 ÷ VIIIU.1	003 VIIIIS IIIAX	

FREQUENCY (with external clamp)						
Range	Resolution	Accuracy	Input voltage	Overload protection		
30.0 ÷ 199.9Hz	0.1Hz	1/0 E0/rda 1 2dat)	1.0~)/ . 1.000)/	605Vrms max		
200 ÷ 400Hz	1Hz	\pm (0.5%rdg+2dgt)	1.0mV ÷ 1.000V	OUSVIINS Max		

Tel.: 03303 / 504066

Fax: 03303 / 504068



M74E

Rel. 1.03 of 19/09/14

Advanced DMM with safety and safety test verifies

Pag 2 of 3

2. TECHNICAL SPECIFICATIONS - SAFETY TEST VERIFY

Continuity test on protective and equalizing conductors						
Range (Ω)	Resolution (Ω)	Accuracy	Overload protection			
0.01 ÷ 19.99	0.01	1/F 00/rdc 1 2dct)	605Vrms max			
20.0 ÷ 99.9	0.1	\pm (5.0%rdg + 3dgt)	005VIIIS IIIAX			

Test current: > 200mA DC for R≤5Ω (included calibration); Resolution on current measurement: 1mA

Open-circuit voltage: $4V \leq V_0 \leq 24V$

Insulation Resistance			
Range (MΩ)	Resolution (M Ω)	Accuracy	Overload protection
0.00 ÷ 19.99	0.01	1/F 00/rdc 1 2dct)	
20.0 ÷ 199.9	0.1	\pm (5.0%rdg + 2dgt)	605Vrms max
200 ÷ 999	1	±(10.0%rdg + 2dgt)	

Test Voltage: 500V DC Test voltage accuracy: -0% ÷ +10% rdg <3.0mA Short circuit current:

Nominal test current: 1mA @ $1k\Omega$ x Vnom ; 1mA @ 500 $k\Omega$

Tripping time test for RC	D type AC and A		
Range (ms)	Resolution (ms)	Accuracy	Overload protection
2 ÷ 300	1	\pm (2.0% rdg + 2dgt)	605Vrms max

30mA, 30x5mA, 100mA, 300mA (Type AC), 30mA (Type A) Nominal trip-out currents:

RCD type: AC (∿), A (♠♠), General

Phase-Earth / Phase-Neutral voltage: 100V ÷ 265V

 $50\text{Hz} \pm 0.5\text{Hz}$ / $60\text{Hz} \pm 0.5\text{Hz}$ Frequency:

Tripping currente test for RCD						
RCD Type	IΔN	Range I∆N [mA]	Resolution	Accuracy		
AC, A (General)	30mA	6.0 ÷ 33.0	0.5mA	- 0%, +10%I _{∆N}		
DI E / DI I	Please Fault / Please Mantaglant towns 400V 200V					

Phase-Earth / Phase-Neutral voltage: 100V ÷ 265V

 $50 Hz \pm 0.5 Hz$ / $60 Hz \pm 0.5 Hz$ Frequency:

Global Earth Resistance						
Test current	Range (Ω)	Resolution (Ω)	Accuracy	Overload protection		
15mA	1 ÷ 1999	1	±(5.0% rdg + 2dgt)	605Vrms max		
100mA	0.1 ÷ 199.9	0.1	±(5.0% rda + 3dat)	005VIIIS IIIAX		

Phase-Earth voltage: 110V ÷ 265V

Frequency: $50 Hz \pm 0.5 Hz$ / $60 Hz \pm 0.5 Hz$

Limit contact voltage: 50V

Phase sequence / conformity			
Type of measure	Voltage range (V)	Frequency range (Hz)	System type
1 Wire	90 ÷ 315 (Phase – Earth)	45 ÷ 65	up to 315 (Phase – Earth) up to 550V (Phase – Phase)
2 Wire	110 ÷ 315 (Phase – Neutral)	45 ÷ 65	up to 315 (Phase – Earth) up to 550V (Phase – Phase)

Max crest factor :1.5

NOTE: the two-wire measurement can be performed also phase to phase in plants without neutral, even with one phase to earth, but always with phase to phase voltage up to 550V



M74E

Rel. 1.03 of 19/09/14

Advanced DMM with safety and safety test verifies

Pag 3 of 3

3. GENERAL SPECIFICATIONS

DISPLAY:

Features: Dual numeric, 9999 points

Display update: 2 times/sec Visible area: 73x73 mm

POWER SUPPLY:

Batteries: 4 batteries 1.5V type LR6-AA-AM3-MN 1500

ELECTRICAL FEATURES:

Conversion: AC 16 Bit, TRMS Sample frequency: 64 sample/period

MECHANICAL FEATURES:

Dimensions (L x W x H): 240 x 100 x 45mm Weight (included batteries): approx 630 g

ENVIRONMENTAL CONDITIONS:

23°C ± 5°C Reference temperature: $0^{\circ} \div 40^{\circ}C$ Working temperature: <70%HR Allowed relative humidity: Storage temperature: -10 ÷ 60°C < 70%HR Storage humidity:

TEST VERIFIES REFERENCE STANDARDS:

Global earth resistance: IEC/EN61557-3 RCDs test: IEC/EN61557-6 Phase sequence indication: IEC/EN61557-7

REFERENCE STANDARDS:

Safety of measuring instruments: IEC/EN61010-1 + A2(1997) IEC/EN61557-1, 3, 6,7 Product type standard: Insulation: double insulation

Pollution degree:

CAT III 550V AC Phase - Ground Overvoltage category: CAT III 550V AC Phase - Phase

Max height of use: 2000m

This instrument complies with the requirements of the European 2006/95/EEC (LVD) and EMC 2004/108/EEC

Tel.: 03303 / 504066

Fax: 03303 / 504068